

Tlv

**Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 9/18/2014**

**Instrument Serial Number: 010587**

**As of 18-Sep-14**

<b>Start Date</b>	<b>End Date</b>	<b>Agency Name</b>
18-Sep-14		DFS Central Lab
15-Apr-14	18-Sep-14	Chase City PD
19-Mar-14	15-Apr-14	DFS Central Lab
01-Apr-13	19-Mar-14	Virginia Beach PD
19-Oct-12	01-Apr-13	DFS Central Lab
18-Dec-08	19-Oct-12	Virginia Peninsula Regional Jail
15-Aug-08	18-Dec-08	DFS Central Lab

TLN

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010587** Worksheet Start Date **9/18/2014**  
Location **Chase City PD**  
Address **219 East 2nd Street Chase City VA 23924**  
DFS Technician **Timmy Neece, Jr** License No. **25387**  
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **746** Reference Barometer (mm HG) **747**  
Reference Barometer(RB)Serial # **009111** RB Calibration Due **5/19/2015**

## Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.295
0.300	0.295	0.286	0.303	Sample 2	0.295
Precision		sample min	sample max	Sample 3	0.294
0.001		0.294	0.295		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.098
0.100	0.098	0.095	0.101	Sample 2	0.098
Precision		sample min	sample max	Sample 3	0.098
0		0.098	0.098		

Dry gas standard Lot No. (with tank no.)

**AG316301-22**

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Instrument Serial Number

010587

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

9/19/14

Issuing Analyst

  
MK

Date

9/19/14

TLN




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>NEECE, JR., TIMMY, L</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>25387</b>	LICENSE EXPIRES <b>10/01/2014</b>	DATE TEST CONDUCTED <b>09/18/2014</b>
TEST EQUIPMENT NUMBER <b>010587</b>		

RESULTS: TIME SAMPLE TAKEN 11:43 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 18 Sep 2014 at 12:10

SLW  
TLW

## Test Results

Instrument Serial Number 010587

Test # 003591 Subject Test

Test Location 1 Department of  
Test Date 18 Sep 2014

Test Location 2 Forensic Science  
Test Time 11:37  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name NEECE, JR.  
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

License Number 25387

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 18 Sep 2014 End Time 11:45

Result Time 11:43

Result Date 18 Sep 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:37

Data Type BLK

Sample Value 0.000

Sample Time 11:38

Data Type CHK

Sample Value 0.098

Sample Time 11:39

Data Type BLK

Sample Value 0.000

Sample Time 11:40

Data Type SUBJ

Sample Value 0.000

Sample Time 11:41

Data Type BLK

Sample Value 0.000

Sample Time 11:42

Data Type SUBJ

Sample Value 0.000

Sample Time 11:43

Data Type BLK

Sample Value 0.000

Sample Time 11:44

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG316301-22

Standard Expiration Date 06/12/2015

Tank Pressure 223

Barometric Pressure 746 mmHg

Blow Sample Number 1 Blow Duration 3.58 sec

Blow Volume 2080 cc End-of-Blow Time 11:41

Blow Sample Number 2 Blow Duration 2.90 sec

Blow Volume 1899 cc End-of-Blow Time 11:43

Tamper Evident Stamp 86cc4f49

Test Status Code 0

Test Status Success

TLN

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010587      Test Number: 3592

Test Date: 09/18/2014      Test Time: 11:47 EDT

Dry Gas Target: 0.098

Lot Number: AG316301-22      Exp Date: 06/12/2015

Tank Pressure: 218 psi      Barometric Pressure: 746 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:48
CHK	0.098	11:48
BLK	0.000	11:50
CHK	0.098	11:51
BLK	0.000	11:52
CHK	0.098	11:53

Test Status: *Success*

Calibration CRC: 4030C8B3

TLN

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010587      Test Number: 3593

Test Date: 09/18/2014      Test Time: 11:57 EDT

Dry Gas Target: 0.295

Lot Number: AG421201-04      Exp Date: 07/31/2016

Tank Pressure: 424 psi      Barometric Pressure: 746 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:57
CHK	0.295	11:58
BLK	0.000	12:00
CHK	0.295	12:00
BLK	0.000	12:02
CHK	0.294	12:03

Test Status: *Success*

Calibration CRC: 4030C8B3